

Guidelines for Use with the Anterior Hip Table.

Ensure the REMUV™ LP is free to inflate and not trapped between tables.



Angle the recovery bed to minimize gaps. Lock and level beds.



During *inflation* of the REMUV™ LP, support the patient by placing your hip against the table. Unless bed rails or securing devices are available, maintain proximity to the patient.



Support the patient's legs at all times during transfer.

